



REPUBLIKA E SHQIPËRISË
MINISTRIA E SHËNDETËSISË
DHE MBROJTJES SOCIALE

CLARIFICATION NO 4

Issued on 21 July 2023

*to the bidding documents for the tender Medical Equipment for Hospitals: Lot 1 Apheresis
Lot 2 Digital mammography Lot 3 Ophthalmology Lot 4 Phaco equipment*

	<i>Request for clarification submitted from potential Bidder</i>	<i>Legal basis</i>	<i>Response to clarification</i>
1	<p><i>LOT 3, ELECTRIC OPERATING BED FOR OPHTHALMIC / ORL /</i></p> <p>Arm movements in the range approx. - 5°/+35° 450 x 600 mm</p> <p><i>The dimension required there is enormous and we think this is not needed at all this big Clinically speaking will not create any benefit. Please accept modification in order to promote broader competition.</i></p> <p><i>Proposed to be changed: Arm movements in the range approx. - 5°/+35° 150 x 500 mm</i></p>	<p>Bidding documents Section technical Specifications</p>	<p><i>LOT 3, ELECTRIC OPERATING BED FOR OPHTHALMIC / ORL /</i></p> <p>Arm movements in the range approx. -5°/+35°</p> <p>Removing the dimensions in order to promote broader competition.</p>